

# HUNTLEIGH

Anwendungshinweise

Kullanım Talimatları

Brugsvejledning

Instrucciones de uso

χρήσης

使用方

Mode d'emploi

Bruksanvisning

Gebruiksaanwijzing

aanwijzing

## INSTRUCTIONS FOR USE

Bruksar

alimatları

使用方法

Käyttöohjeet

Instruções de Utilização

Istruzioni per l'uso

Anwendungshinweise

Οδηγίες χρήσης

Anwendungshinweise

### Hydroven 3

# Contents

|                                                          |           |
|----------------------------------------------------------|-----------|
| <b>1. Safety .....</b>                                   | <b>3</b>  |
| 1.1 Warnings.....                                        | 3         |
| <b>2. Electromagnetic Compatibility .....</b>            | <b>5</b>  |
| <b>3. Introduction.....</b>                              | <b>8</b>  |
| 3.1 About this Manual.....                               | 8         |
| 3.2 Intended Use .....                                   | 8         |
| 3.3 About the Hydroven 3 system .....                    | 8         |
| 3.4 Use Environment .....                                | 8         |
| <b>4. Clinical Applications .....</b>                    | <b>9</b>  |
| 4.1 Indications.....                                     | 9         |
| 4.2 Contraindications .....                              | 10        |
| <b>5. Preliminary Checks .....</b>                       | <b>11</b> |
| <b>6. Clinical Treatment Guide .....</b>                 | <b>12</b> |
| <b>7. Garment and Insert Information.....</b>            | <b>13</b> |
| 7.1 Garment Description .....                            | 13        |
| 7.2 Selecting the correct Garment.....                   | 13        |
| 7.3 Applying the Garment .....                           | 14        |
| <b>8. Operation .....</b>                                | <b>16</b> |
| 8.1 Pump Description .....                               | 16        |
| 8.2 Operation.....                                       | 17        |
| 8.2.1 To Adjust the Pressure Control Knob Position ..... | 17        |
| 8.2.2 Switch On .....                                    | 17        |
| 8.2.3 To Set the Garment Pressure .....                  | 18        |
| 8.2.4 Shut Down .....                                    | 18        |
| 8.2.5 To Remove the Garment.....                         | 18        |
| <b>9. Decontamination .....</b>                          | <b>19</b> |
| 9.1 Cleaning.....                                        | 19        |
| 9.2 Chemical Disinfection .....                          | 19        |
| 9.3 Cleaning and Sterilising Garments .....              | 19        |

|                                                      |           |
|------------------------------------------------------|-----------|
| <b>10.Routine Maintenance .....</b>                  | <b>20</b> |
| <b>10.1 Hydroven 3 System.....</b>                   | <b>20</b> |
| 10.1.1 Maintenance .....                             | 20        |
| 10.1.2 Servicing .....                               | 20        |
| 10.1.3 Service Period.....                           | 20        |
| <b>10.2 Hydroven 3 Pump .....</b>                    | <b>20</b> |
| 10.2.1 General Care, Maintenance and Inspection..... | 20        |
| 10.2.2 Serial Labels .....                           | 20        |
| <b>11.Trouble Shooting .....</b>                     | <b>21</b> |
| <b>12.Accessories .....</b>                          | <b>22</b> |
| <b>13.Specifications .....</b>                       | <b>24</b> |
| <b>13.1 Equipment Classification .....</b>           | <b>24</b> |
| <b>13.2 General .....</b>                            | <b>24</b> |
| <b>13.3 Environmental.....</b>                       | <b>25</b> |
| <b>13.4 Standards Compliance.....</b>                | <b>25</b> |
| <b>14.Product Labelling .....</b>                    | <b>26</b> |
| <b>15.End of Life Disposal .....</b>                 | <b>28</b> |
| <b>16.Warranty &amp; Service.....</b>                | <b>29</b> |
| <b>16.1 Service Returns.....</b>                     | <b>29</b> |

# 1. Safety



***Before using this equipment, please study this manual carefully and familiarise yourself with the controls, display features and operation. Ensure that each user fully understands the safety and operation of the unit, as misuse may cause harm to the user or patient, or damage to the product.***

***Please keep these Instructions for Use to hand for future reference.***

## Symbols



General Warning



Follow Instructions for Use

## 1.1 Warnings

Do NOT use unapproved accessories or attempt to modify, disassemble or otherwise misuse the Hydroven 3 system. Failure to observe this caution could result in injury, or in extreme cases, death.



***A possible explosion hazard exists if used in the presence of flammable anaesthetics.***



***Do not operate the unit from the mains supply if the mains cable is damaged.***



***Do not immerse any portion of the unit in water or other liquids.***



***Use only recommended accessories listed in this manual.***



***If this product is connected to another item of electrical equipment, it is important that the system is fully compliant with EN60601-1.***



***It is the responsibility of the care giver to ensure that the user can use this product safely.***



***Make sure that the mains power cable and tubeset or air hoses are positioned to avoid causing a trip or other hazard, and are clear of moving bed mechanisms or other possible entrapment areas.***



***Electrical equipment may be hazardous if misused. There are no user-serviceable parts inside the pump. The pump's case must only be removed by authorised technical personnel. No modification of this equipment is allowed.***



***The mains power socket/plug must be accessible at all times. To disconnect the pump completely from the electricity supply, remove the plug from the mains power socket.***



***Disconnect the pump from the mains power socket before cleaning and inspecting.***



***Only the pump and garment/insert combination as indicated by Huntleigh should be used. The correct function of the product cannot be guaranteed if incorrect pump and garment combinations are used.***



***Bags supplied with this equipment may present a suffocation risk; to avoid the risk of suffocation keep the bags away from babies and small children.***



***Caution (applicable to the USA market only)  
US Federal law restricts this device to sale by or on the order of a physician.***



***Do not expose the system to naked flames, such as cigarettes, etc.***



***Do not store the system in direct sunlight.***



***Do not use phenol-based solutions to clean the system.***



***Make sure the system is clean and dry prior to use or storage.***



***Pets and children must be supervised in the vicinity of the system.***

## Expected Service Life

The Hydroven 3 has an expected service life of 7 years. To maintain the condition of the pump have it serviced regularly according to the schedule recommended by your Huntleigh distributor.

## 2. Electromagnetic Compatibility

Make sure the environment in which Hydroven 3 is installed is not subject to strong sources of electromagnetic interference (e.g. radio transmitters, mobile phones).

This equipment generates and uses radio frequency energy. If not installed and used properly, in strict accordance with the manufacturer's instructions, it may cause or be subject to interference. Type-tested in a fully configured system, complies with EN60601-1-2, the standard intended to provide reasonable protection against such interference. Whether the equipment causes interference may be determined by turning the equipment off and on. If it does cause or is affected by interference, one or more of the following measures may correct the interference:

- Reorienting the equipment
- Relocating the equipment with respect to the source of interference
- Moving the equipment away from the device with which it is interfering
- Plugging the equipment into a different outlet so that the devices are on different branch circuits



***If this equipment needs to be used adjacent to other electrical equipment, normal operation must be checked before use.***

| Guidance and Manufacturer's declaration - electromagnetic emissions                                                                                                                     |            |                                                                                                                                                                                                                             |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| The Hydroven 3 is intended for use in the electromagnetic environment specified below. The customer or the user of the Hydroven 3 should assure that it is used in such an environment. |            |                                                                                                                                                                                                                             |
| Emissions Test                                                                                                                                                                          | Compliance | Electromagnetic Environment - guidance                                                                                                                                                                                      |
| RF emissions<br>CISPR 11                                                                                                                                                                | Group 1    | The Hydroven 3 uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.                                         |
| RF emissions<br>CISPR 11                                                                                                                                                                | Class B    |                                                                                                                                                                                                                             |
| Harmonic emissions<br>IEC 61000-3-2                                                                                                                                                     | Class A    |                                                                                                                                                                                                                             |
| Voltage fluctuations /<br>flicker emissions<br>IEC 61000-3-3                                                                                                                            | Complies   | The Hydroven 3 is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes. |

| Guidance and Manufacturer's declaration - electromagnetic immunity                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |                           |                  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------|------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| The Hydroven 3 is intended for use in the electromagnetic environment specified below. The customer or the user of the Hydroven 3 should assure that it is used in such an environment.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          |                           |                  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |
| Immunity Test                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    | IEC 60601 test level      | Compliance level | Electromagnetic Environment - guidance                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |                           |                  | Portable and mobile RF communications equipment should be used no closer to any part of the Hydroven 3, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.                                                                                                                                                                                                                                                                                                                                                  |
| Conducted RF<br>IEC 61000-4-6                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    | 3 Vrms<br>150kHz to 80MHz | 3V               | $d = 1.2 \sqrt{P}$                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
| Radiated RF<br>IEC 61000-4-3                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     | 3 Vrms<br>80MHz to 2.5MHz | 3V/m             | $d = 1.2 \sqrt{P}$ 80MHz to 800MHz<br>$d = 2.3 \sqrt{P}$ 800MHz to 2.5GHz                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |
|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |                           |                  | <p>where <math>P</math> is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and <math>d</math> is the recommended separation distance in metres (m).</p> <p>Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, <sup>a</sup> should be less than the compliance level in each frequency range <sup>b</sup>. Interference may occur in the vicinity of the equipment marked with the following symbol:</p>  |
| <p>NOTE 1 At 80MHz and 800MHz, the higher frequency range applies.</p> <p>NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.</p>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |                           |                  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |
| <p><sup>a</sup> Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the Hydroven 3 is used exceeds the applicable RF compliance level above, the Hydroven 3 should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the Hydroven 3.</p> <p><sup>b</sup> Over the frequency range 150kHz to 80kHz, field strengths should be less than 3V/m.</p> |                           |                  |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |

| Recommended separation distances between portable and mobile RF communications equipment and the Hydroven 3                                                                                                                                                                                                                                                                                                                          |                                                                |                                           |                                            |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------|-------------------------------------------|--------------------------------------------|
| The Hydroven 3 is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or user of the Hydroven 3 can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Hydroven 3 as recommended below, according to the maximum output power of the communications equipment. |                                                                |                                           |                                            |
| Rated maximum output power of transmitter<br><br>W                                                                                                                                                                                                                                                                                                                                                                                   | Separation distance according to frequency of transmitter<br>m |                                           |                                            |
|                                                                                                                                                                                                                                                                                                                                                                                                                                      | 150kHz to 80MHz<br><br>$d = 1.2 \sqrt{P}$                      | 80MHz to 800MHz<br><br>$d = 1.2 \sqrt{P}$ | 800MHz to 2.5GHz<br><br>$d = 2.3 \sqrt{P}$ |
| 0.01                                                                                                                                                                                                                                                                                                                                                                                                                                 | 0.12                                                           | 0.12                                      | 0.23                                       |
| 0.1                                                                                                                                                                                                                                                                                                                                                                                                                                  | 0.38                                                           | 0.38                                      | 0.73                                       |
| 1                                                                                                                                                                                                                                                                                                                                                                                                                                    | 1.2                                                            | 1.2                                       | 2.3                                        |
| 10                                                                                                                                                                                                                                                                                                                                                                                                                                   | 3.8                                                            | 3.8                                       | 7.3                                        |
| 100                                                                                                                                                                                                                                                                                                                                                                                                                                  | 12                                                             | 12                                        | 23                                         |
| For transmitters rated at a maximum output power not listed above, the recommended separation distance $d$ in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where $P$ is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.                                                                                                 |                                                                |                                           |                                            |
| NOTE 1 At 80MHz and 800MHz, the separation distance for the higher frequency range applies.                                                                                                                                                                                                                                                                                                                                          |                                                                |                                           |                                            |
| NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.                                                                                                                                                                                                                                                                   |                                                                |                                           |                                            |

## 3. Introduction

### 3.1 About this Manual

This manual is your introduction to the Hydroven® 3 system.

You must read and fully understand this manual before using the system. Use this manual to initially set up the system, and keep it as a reference for day-to-day routines and as a guide to maintenance.

If you have any difficulties in setting-up or using the Hydroven 3 system, contact your local Huntleigh sales representative, listed at the end of this manual.

### 3.2 Intended Use

The intended use of this product is to manage the list of clinical conditions detailed in the “Indications” section.

The Hydroven 3 system should be used as part of a prescribed plan of care detailed in the “Indications” section.

### 3.3 About the Hydroven 3 system

The pump supplies air via connecting tubes to an inflatable garment allowing the application of controlled pressure to gently compress the limb. This action assists in increasing the return of blood, excess fluids, improves venous stasis and encourages the reabsorption of waste products.

The pump operates on an automatically timed cycle of 3 minutes, 90 seconds inflation followed by 90 seconds deflation. Variable pressure output ranges from 30-100 mmHg. The garments are inflated alternately.

The Hydroven 3 system operates two types of garments:

- Hydroven 1 garments have a single chamber and provide uniform compression.
- Hydroven 3 garments have three chambers providing graduated segmental compression, inflating distally to proximally.

Optional garment inserts can be used to increase the circumference of the standard arm and leg garments.

A full technical description of the Hydroven 3 system can be found in the Service Manual, part No. SER0014, available from your local Huntleigh sales office.

### 3.4 Use Environment

Hydroven 3 is suitable for use in hospital, primary care and community settings. It must not be used outdoors, or in any environment where it may come into contact with water.

## 4. Clinical Applications

### 4.1 Indications

Intermittent Pneumatic Compression (IPC) is effective in the treatment of the following clinical conditions, when combined with an individualised monitoring programme:

- Oedema.
  - Dependent (including secondary to cerebrovascular incident, pregnancy or paralysis).
  - Traumatic (post-surgical or injury).
- Lymphoedema.
  - Primary and secondary (including post surgery, radio or chemotherapy).
- Chronic venous insufficiency.
- Post phlebotic syndrome.
- Acute and chronic wounds including venous leg ulcers and post-surgical wounds.

IPC may also be beneficial in the management of:

- Fixed flexion deformity.
- Arthritic conditions.
- Lower limb pain due to trauma or surgery.
- Lipoedema.

Selection should be based upon a holistic assessment of the patients' individual care needs.

**Note:** *These systems represent one aspect of a treatment strategy; if the patient's condition changes the overall therapy regimen should be reviewed by the prescribing clinician.*

**Note:** *The above are guidelines only and should not replace clinical judgement.*

## 4.2 Contraindications

IPC should NOT be used in the following circumstances:

- Known or suspected deep vein thrombosis (DVT), pulmonary embolism, thrombophlebitis and acute infections of the skin, such as cellulitis.
- Decompensated/severe congestive cardiac failure, pulmonary oedema associated with significant limb oedema or any condition where an increase of fluid to the heart may be detrimental.
- Severe arteriosclerosis or other ischaemic vascular disease.
- Active metastatic disease affecting the limb.

**NOTE TO PATIENT:** *if you are uncertain whether you have any of the above conditions please consult a physician before use.*



**CAUTION:** *IPC should be used with care in patients with the following symptoms or conditions:*

- *Peripheral neuropathy, pain or numbness in the limb.*
- *Undiagnosed, untreated or infected wounds, fragile skin, grafts or dermatological conditions that may be aggravated by the garment.*
- *Extreme limb deformity which may practically impede the correct application of the garment.*



**WARNING:** *Therapy should be interrupted if pain, tingling or numbness of the limb occurs during or as a result of therapy.*



**WARNING:** *In the event of a power failure or fault whereby the garment remains inflated, remove the garment(s) from the patient's limbs.*



**WARNING:** *Patients must not walk or stand when wearing leg garments.*

## 5. Preliminary Checks

### *Contents (supplied with each system)*

| Item           | Item                     |
|----------------|--------------------------|
| 1 x Hydroven 3 | 1 x Instructions for Use |

### ***Delivery Inspection***

Huntleigh Healthcare Ltd takes every precaution to ensure that goods reach you in perfect condition. However, accidental damage can occur in transit and storage. For this reason we recommend that a thorough visual inspection is made immediately the unit is received. Should any damage be evident or any parts missing, ensure that Huntleigh Healthcare Ltd is informed at once.

### ***Storage***

If the unit not be required for immediate use, it should be re-sealed into its original packing after carrying out the initial delivery inspection, and stored under covered conditions at a temperature between -20°C to +50°C, and relative humidity of 20% to 95% non-condensing.

After exposure to extreme temperatures during storage, the pump must be allowed to adjust to normal operating temperatures for a minimum of 12 hours before use. Failure to do this may result in accelerated wear of mechanical components.

## 6. Clinical Treatment Guide

An initial pressure setting of 40 mmHg is suggested at the commencement of treatment. It may be necessary to start at a lower level of pressure, dependent on the patient's tolerance.

The pressure can be gradually increased over time, until the required pressure is reached. The upper treatment pressure range is generally 60-70 mmHg.


A single treatment session is usually 20-30 minutes.

**Note:**     *The above settings and timings are guidelines, and should not be used as a substitute for clinical judgement and experience.*


**Note:**     *Loss of mains power will halt therapy.*

## 7. Garment and Insert Information

### 7.1 Garment Description



Hydroven 1 Garment




Hydroven 3 Garment


**Note:** For Hydroven 1 garments, when an insert is not in use, the bung can be used as a quick air release.

### 7.2 Selecting the correct Garment

1. Select the type of garment depending on treatment type:
  - Hydroven 1 garments have a single chamber and provide uniform compression.
  - Hydroven 3 garments have three chambers providing graduated segmental compression, inflating distally to proximally.
2. Measure the circumference of the largest part of the limb, and the length in cm/inch from the heel to the upper thigh for a full leg garment, heel to knee for half leg garment, from shoulder to finger tips for full arm garment, and from elbow to finger tips for half arm garment. Refer to "Accessories" section to order the correct size garment.



Half Leg Garment




Full Arm Garment

## 7.3 Applying the Garment



**WARNING:** Bags supplied with this equipment may present a suffocation risk; to avoid the risk of suffocation, keep the bags away from babies and small children.




**Note:** Before fitting the garment ensure all quick air release bungs are closed, as this will effect the efficiency of the garment.

1. If a larger circumference is required, fit a matching length insert piece before applying to the limb. If appropriate, a primary dressing or stockinette may be used underneath the garments.
2. Undo the zip on the garment.
3. If a garment insert is fitted to the garment, fully fasten one of the zips between the garment and insert, leaving the other unfastened.
4. Before applying the garment (and insert, if fitted) to the limb, zip up the first 150 mm (6") of the unfastened garment zip. Put the garment (and insert) onto the limb and fully fasten the zip. Make sure that the quick air release bung is secured.
5. Make sure the patient is in a comfortable position with the limb supported or elevated as necessary.
6. Check that the insert piece connecting tube is not kinked and is attached to the garment using the lower outlet bung as shown below.
7. Attach the garment tubing to the pump ensuring a "click" is heard from each snap-lock connector.
8. If only one garment is to be used, attach the garment to either port on the pump. The system will automatically identify that only one garment is to be used.

**Note:** Ensure that the all zips are fully done up on the garment before switching the pump on.

9. Switch on the pump and adjust the pressure control accordingly.



**CAUTION:** Do not apply the garment to the limb unless it is partially zipped, as you may damage the garment zip.




**CAUTION:** Do not apply or remove the garment while it is attached to the pump and the pump is in operation, as you may damage the garment zip.



**CAUTION:** Do not stand or walk while leg garments are fitted.

## 8. Operation

### 8.1 Pump Description



| Item No. | Description                                        | Function                                                                                                                                                                                                                                                                                  |
|----------|----------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1        | On/Off Switch                                      | Operation of this switch Starts or Stops the system                                                                                                                                                                                                                                       |
| 2        | Pressure Control Knob / Lock Pin*<br>(* If fitted) | Rotate clockwise to increase pressure or counter-clockwise to decrease pressure (pressure range 30 ~ 100 mmHg)<br><br>The pressure control knob is locked in position, (if Lock Pin is fitted), to prevent accidental movement. Refer to "To Adjust the Pressure Control Knob Position" . |
| 3        | Tube Connectors                                    | Snap-lock connectors for Garment attachment                                                                                                                                                                                                                                               |
| 4        | Pressure Gauge                                     | Indicates delivery pressure to garment                                                                                                                                                                                                                                                    |
| 5        | Carry Handle                                       | For easy handling of the pump                                                                                                                                                                                                                                                             |

**Note:** *If the operation of performance of the pump changes during use, refer to "Trouble Shooting" section of this IFU before calling a service engineer or contacting your local Huntleigh sales office.*



**CAUTION:** Do not apply or remove the garment while it is attached to the pump and the pump is in operation, as you may damage the garment zip.

## 8.2 Operation



***It is the responsibility of the care giver to ensure that the user can use this product safely.***


The pump should be placed securely on a flat surface.

Before starting the pump ensure that the garments are properly applied, the zippers are secured and the garment connecting tubes are attached to the pump outlet ports via the snap-lock connectors (3).

### 8.2.1 To Adjust the Pressure Control Knob Position

The pressure control knob (2) is locked in position\* to prevent accidental rotation.

To adjust the position of the pressure control knob:



1. Lift the lock pin\* (A) to release the control knob.
2. Rotate the control knob (B) whilst the lock pin is raised.
3. Release the lock pin\* when the pressure control knob is in the desired position to lock the control knob.

*\* If lock pin fitted.*

**Note:** ***Rotate the pressure control knob clockwise to increase and counterclockwise to decrease pressure.***

Make sure that the pressure control knob is set to minimum i.e. rotated fully counterclockwise.



***WARNING: Ensure that all cables and air hoses are positioned so that they do not present a trip hazard or strangulation.***



***WARNING: Bags supplied with this equipment may present a suffocation risk; to avoid the risk of suffocation keep the bags away from babies and small children.***

### 8.2.2 Switch On

Connect the pump to the mains power supply using the power cable provided. Turn the mains power switch (1) to the On (I) position.

### 8.2.3 To Set the Garment Pressure

While the garment is inflating, rotate the pressure control knob (2) slowly clockwise until the required pressure is displayed on the gauge (4). The garments will take approximately three cycles to fully inflate. Check and adjust as necessary after three inflation cycles.

**Note:** *It might be necessary to start at a lower pressure level dependant on the patient's tolerance. Compression should not cause any discomfort or pain to the patient.*

### 8.2.4 Shut Down

Turn the power switch (1) to the off (O) position. Turning the power off will stop the patient therapy.

**Note:** *If it is required to completely isolate the pump from the mains power, remove the plug from the mains power socket.*

### 8.2.5 To Remove the Garment

Make sure the pump power switch is in the off (O) position, disconnect the tubing from the pump by removing the snap-lock connectors (3), and release the quick air release cap on the garment.

Only open the zip after the garment is completely deflated.

## 9. Decontamination

The following processes are recommended, but should be adapted to comply with the local or national guidelines (Decontamination of Medical Devices) which may apply within the Healthcare Facility or the country of use. If you are uncertain, you should seek advice from your local Infection Control Specialist.

The Hydroven 3 system should be routinely decontaminated between patients and at regular intervals while in use; as is good practice for all reusable medical devices.



**WARNING:** Remove the electrical supply to the pump by disconnecting the mains power cord from the mains power supply before cleaning.  
Protective clothing should always be worn when carrying out decontamination procedures.



**CAUTION:** Do not use Phenol-based solutions or abrasive compounds or pads during the decontamination process as these will damage the surface coating. Avoid immersing electrical parts in water during the cleaning process. Do not spray cleaning solutions directly onto the pump. Do not immerse the tubeset in water.

### 9.1 Cleaning

Clean all exposed surfaces and remove any organic debris by wiping with a cloth moistened with a simple (neutral) detergent and water.  
Do not allow water or cleaning solutions to collect on the surface of the pump.

### 9.2 Chemical Disinfection

We recommend a chlorine-releasing agent, such as sodium hypochlorite, at a strength of 1,000ppm available chlorine (this may vary from 250ppm to 10,000ppm depending on local policy and contamination status).

Wipe all cleaned surfaces with the solution, then wipe using a cloth moistened with water and dry thoroughly.

Alcohol based disinfectants (strength 70%) may be used as an alternative.

Ensure the product is dry before storage.

If an alternative disinfectant is selected from the wide variety available we recommend that suitability for use is confirmed with the chemical supplier prior to use.

### 9.3 Cleaning and Sterilising Garments

Wipe down garments using a neutral detergent or soap powder at 40°C. Dry thoroughly.

After cleaning gas sterilisation is possible, however:

- Do not exceed 51°C (120°F).
- Do not autoclave.

## **10. Routine Maintenance**

### **10.1 Hydroven 3 System**

#### **10.1.1 Maintenance**

The equipment has been designed to be maintenance-free between service periods.

#### **10.1.2 Servicing**

Huntleigh will make available on request service manuals, component parts lists and other information necessary for Huntleigh trained personnel to repair the system.

#### **10.1.3 Service Period**

Huntleigh recommend that the Hydroven 3 pump is serviced every 12 months by a Huntleigh authorised service agent.

### **10.2 Hydroven 3 Pump**

#### **10.2.1 General Care, Maintenance and Inspection**

Check all electrical connections and power cable for signs of excessive wear.

Check the tubeset and connectors for any damage.

In the event of the pump being subjected to abnormal treatment, e.g. immersed in water or dropped, the unit must be returned to an authorised service centre.

#### **10.2.2 Serial Labels**

The serial number for the pump is on the label on the back of the pump case. Quote this serial number when requesting service.

# 11. Trouble Shooting

If you should encounter a problem, please follow the fault finding guide below. If the fault cannot be rectified, please refer to Service.

| Fault                                       | Check                                 | Remedy                                  |
|---------------------------------------------|---------------------------------------|-----------------------------------------|
| Pump does not operate.                      | Is power switch on?                   | Check switch.                           |
|                                             | Is power cord plugged in correctly?   | Check connections.                      |
|                                             | Fuse blown?                           | Call service engineer.                  |
| Pump operates but garment will not inflate. | Blockage in garment supply tube.      | Ensure that the tube airway is clear.   |
|                                             | Garment not fitted correctly to pump. | Check connections.                      |
|                                             | Pressure control set too low.         | Increase pressure control.              |
|                                             | Air leak in garment.                  | Check garment.<br>Replace if defective. |

**Note:** *If the trouble shooting procedures do not return the system to normal performance, stop using the system immediately and call the service engineer or return the unit to Huntleigh for service. Refer to “Warranty & Service” .*

## 12. Accessories



**WARNING:** Use only recommended accessories listed in this manual.

### Garments

| Hydroven 1 Leg Garment |          |            |               |
|------------------------|----------|------------|---------------|
| Order Code             | Type     | Length (L) | Circumference |
| 5101L50                | Half Leg | 50 cm      | 61 cm         |
| 5101L66                | Full Leg | 66 cm      | 64 cm         |
| 5101L71                | Full Leg | 71 cm      | 66 cm         |
| 5101L76                | Full Leg | 76 cm      | 72 cm         |
| 5101L84                | Full Leg | 84 cm      | 72 cm         |
| 5101L92                | Full Leg | 92 cm      | 72 cm         |

| Hydroven 1 Arm Garment |          |            |                           |                               |
|------------------------|----------|------------|---------------------------|-------------------------------|
| Order Code             | Type     | Length (L) | Circumference at hand (H) | Circumference at shoulder (S) |
| 5101A51                | Half Arm | 51 cm      | 44 cm                     | 56 cm                         |
| 5101A68                | Full Arm | 68 cm      | 44 cm                     | 62 cm                         |
| 5101A78                | Full Arm | 78 cm      | 44 cm                     | 62 cm                         |

| Hydroven 3 Leg Garment |          |            |               |
|------------------------|----------|------------|---------------|
| Order Code             | Type     | Length (L) | Circumference |
| 5103L50                | Half Leg | 50 cm      | 61 cm         |
| 5103L66                | Full Leg | 66 cm      | 64 cm         |
| 5103L71                | Full Leg | 71 cm      | 66 cm         |
| 5103L76                | Full Leg | 76 cm      | 72 cm         |
| 5103L84                | Full Leg | 84 cm      | 72 cm         |
| 5103L92                | Full Leg | 92 cm      | 72 cm         |

| Hydroven 3 Arm Garment |          |            |                           |                               |
|------------------------|----------|------------|---------------------------|-------------------------------|
| Order Code             | Type     | Length (L) | Circumference at hand (H) | Circumference at shoulder (S) |
| 5103A68                | Full Arm | 68 cm      | 44 cm                     | 62 cm                         |
| 5103A78                | Full Arm | 78 cm      | 44 cm                     | 62 cm                         |

*Inserts*

| Hydroven Garment Insert Pieces<br>(to fit Hydroven 1 and 3 Garments) |          |            |                           |                             |
|----------------------------------------------------------------------|----------|------------|---------------------------|-----------------------------|
| Order Code                                                           | Type     | Length (L) | Circumference<br>Wide End | Circumference<br>Narrow End |
| 510LI50                                                              | Half Leg | 50 cm      | 19 cm                     | 14 cm                       |
| 510LI66                                                              | Full Leg | 66 cm      | 19 cm                     | 14 cm                       |
| 510LI71                                                              | Full Leg | 71 cm      | 19 cm                     | 14 cm                       |
| 510LI76                                                              | Full Leg | 76 cm      | 19 cm                     | 14 cm                       |
| 510LI84                                                              | Full Leg | 84 cm      | 19 cm                     | 14 cm                       |
| 510LI92                                                              | Full Leg | 92 cm      | 19 cm                     | 14 cm                       |
| 510AI68                                                              | Full Arm | 68 cm      | 17 cm                     | 12 cm                       |
| 510AI78                                                              | Full Arm | 78 cm      | 17 cm                     | 12 cm                       |

# 13. Specifications

## 13.1 Equipment Classification

|                                                                                                                                  |                                                                                                                                               |
|----------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------|
| Type of protection against electric shock.                                                                                       | Class II, Double Insulated                                                                                                                    |
| Degree of protection against electric shock<br> | Type BF                                                                                                                                       |
| Mode of operation.                                                                                                               | Continuous                                                                                                                                    |
| Degree of protection against solid and liquid ingress                                                                            | IP21* - Protection against ingress of solid objects more than 12.5mm diameter and water droplets falling vertically.<br>IPX0* - No Protection |
| Degree of safety of application in the presence of a flammable anaesthetic                                                       | Equipment not suitable for use in the presence of a FLAMMABLE ANAESTHETIC MIXTURE WITH AIR, OXYGEN OR NITROUS OXIDE                           |

## 13.2 General

|                  |                                                                                           |
|------------------|-------------------------------------------------------------------------------------------|
| Model            | Hydroven 3                                                                                |
| Part Numbers     | 510001      UK<br>510009AU    Australia & New Zealand<br>510009ZA    South Africa & India |
| Pressure Range   | 30 - 100 mmHg $\pm$ 5%                                                                    |
| Supply voltage   | 230 V AC                                                                                  |
| Supply Frequency | 50Hz                                                                                      |
| Pump Fuse Rating | F500 mAH 250 V                                                                            |
| Plug Fuse Rating | 5A to BS1362 (UK ONLY)                                                                    |
| Power input      | 14 VA                                                                                     |
| Case Material    | Fire Retardent ABS Plastic                                                                |
| Size             | 270 x 130 x 150 mm (10.6 x 5.1 x 5.9")                                                    |
| Weight           | 2.5 kg (5.5 lb)                                                                           |

\* See product label for IP Rating

### 13.3 Environmental

| Condition                             | Temperature range                 | Relative Humidity              | Atmospheric Pressure |
|---------------------------------------|-----------------------------------|--------------------------------|----------------------|
| Operating                             | 5°C to 40°C<br>(41°F to 104°F)    | 30% to 75%<br>(non condensing) | 700 to 1060 hPa      |
| Storage and transport<br>(Long term)  | 10°C to 40°C<br>(50°F to 104°F)   | 20% to 95%<br>(non condensing) | 700 to 1060 hPa      |
| Storage and transport<br>(short term) | -25°C to 70°C<br>(-13°F to 158°F) | 20% to 95%                     | 500 to 1060 hPa      |

**Note:** *When exposed to extreme temperature during storage, the pump must be allowed to adjust to normal temperatures for a minimum of 12 hours before use. Failure to do so may result in accelerated wear of mechanical components.*


### 13.4 Standards Compliance

|                                                             |
|-------------------------------------------------------------|
| EN60601-1:1990/A13:1996 and IEC 60601-1:1988/A2:1995        |
| EN60601-1-1:2001 and EN60601-1-2: 2001                      |
| UL60601-1, UL2601-1 and CAN/CSA C22.2 No. 601.1-M90         |
| EN60601-1:2006, EN60601-1-11:2010* and IEC 60601-1:2005     |
| AAMI/ANSI ES60601-1:2006 and CAN/CSA C22.2 No.60601.1(2008) |
| EN62366:2008                                                |
| BS EN 980:2008                                              |


*\* Only applies to IP21 rated products (see product label for IP rating)*

# 14. Product Labelling

| Symbols                                                                             |                                                                                                                                                                                                                                              |                                                                                     |                                         |
|-------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------|-----------------------------------------|
|    | Hydroven 3 is Class II, double insulated according to the definitions in BS EN 60601-1:1990                                                                                                                                                  |                                                                                     |                                         |
|    | Applied parts are type BF according to the definitions in BS EN 60601-1:1990                                                                                                                                                                 |                                                                                     |                                         |
|    | Refer to this document (Instructions for Use) for a description of the product classification (3rd Edition).                                                                                                                                 |                                                                                     |                                         |
|    | Alternating Current (AC)                                                                                                                                                                                                                     |                                                                                     |                                         |
|   | With respect to electric shock, fire and mechanical hazards only in accordance with CAN/CSA-C22.2 No. 60601.1 (2008).<br><b>MEDICAL EQUIPMENT</b>                                                                                            |                                                                                     |                                         |
|  | This symbol signifies that this product, including its accessories and consumables is subject to the WEEE (Waste Electrical and Electronic Equipment) regulations and should be disposed of responsibly in accordance with local procedures. |                                                                                     |                                         |
|  | Refer to this document (Instructions for Use) for a description of the product classification (2nd Edition).                                                                                                                                 |                                                                                     |                                         |
|  | This symbol signifies that this product complies with the essential requirements of the Medical Devices Directive 93/42/EEC as amended by 2007/47/EC                                                                                         |                                                                                     |                                         |
|  | Power:<br>Disconnects from the<br>mains supply                                                                                                                                                                                               |  | Power:<br>Connects to the mains supply  |
|  | Follow Instructions for<br>Use                                                                                                                                                                                                               |  | Fuse                                    |
|  | Serial Number                                                                                                                                                                                                                                |  | Reference Number                        |
|  | Manufacturer                                                                                                                                                                                                                                 |  | Cardboard packaging can be<br>recycled. |

| Cleaning Symbols                                                                  |                              |                                                                                   |                                                        |
|-----------------------------------------------------------------------------------|------------------------------|-----------------------------------------------------------------------------------|--------------------------------------------------------|
|  | Wipe surface with damp cloth |  | Use solution diluted to 1000 ppm of Available Chlorine |
|  | Do not iron                  |  | Do Not Use Phenol-based cleaning Solutions             |
|  | Do not dry clean             |  | Do not tumble dry                                      |

## 15. End of Life Disposal



This symbol signifies that this product, including its accessories and consumables is subject to the WEEE (Waste Electrical and Electronic Equipment) regulations and should be disposed of responsibly in accordance with local procedures.

## 16. Warranty & Service

Huntleigh Healthcare Diagnostic Products Division standard terms and conditions apply to all sales. A copy is available on request. These contain full details of warranty terms and do not limit the statutory rights of the consumer.

### 16.1 Service Returns

#### United Kingdom only

If for any reason the Hydroven 3 has to be returned, please:

- Clean the product following the instructions in this manual.
- Pack it in suitable packing.
- Attach a decontamination certificate (or other statement declaring that the product has been cleaned) to the outside of the package.
- Mark the package 'Service Department '

For further details, refer to NHS document HSG(93)26.

Huntleigh Healthcare Ltd reserve the right to return product that does not contain a decontamination certificate.

Service Department.  
Huntleigh Healthcare, Diagnostic Products Division,  
35, Portmanmoor Rd.,  
Cardiff. CF24 5HN  
United Kingdom.

Tel: +44 (0)29 20485885  
Fax: +44 (0)29 20492520  
Email: [sales@huntleigh-diagnostics.co.uk](mailto:sales@huntleigh-diagnostics.co.uk)  
[service@huntleigh-diagnostics.co.uk](mailto:service@huntleigh-diagnostics.co.uk)  
[www.huntleigh-diagnostics.com](http://www.huntleigh-diagnostics.com)

#### International Customers

Contact your local office listed on the inside back cover.

**INTERNATIONAL**

ArjoHuntleigh International Ltd  
ArjoHuntleigh House  
Houghton Hall Park  
Houghton Regis  
UK-DUNSTABLE LU5 5XF  
Tel: +44 (0) 1582 745 800  
Fax: +44 (0) 1582 745 866  
E-mail: international@arjohuntleigh.com

**DEUTSCHLAND**

HNE Healthcare GmbH  
Industriering Ost 66  
DE-47906 Kempen  
Tel: +49 (0) 2152 551110  
Fax: +49 (0) 2152 551120  
E-mail: verkauf@huntleigh.de

**SVERIGE**

ARJO Scandinavia AB  
Hans Michelsensgatan 10  
SE-211 20 MALMÖ  
Tel: +46 (0) 10 494 7760  
Fax: +46 (0) 10 494 7761  
E-mail: kundservice@arjohuntleigh.com

**NORGE**

ArjoHuntleigh Norway AS  
Olaf Helsets vei 5  
N-0694 OSLO  
Tel: +47 22 08 00 50  
Faks: +47 22 08 00 51  
E-mail: no.kundeservice@arjohuntleigh.com

**DANMARK**

ArjoHuntleigh A/S  
Vassingerødvej 52  
DK-3540 LYNGE  
Tel: +45 49 13 84 86  
Fax: +45 49 13 84 87  
E-mail: dk\_kundeservice@arjohuntleigh.com

**USA**

ArjoHuntleigh Inc.  
2349 W Lake Street Suite 250  
US-Addison, IL 60101  
Tel: +1 630 307 2756  
Free: +1 800 323 1245 Institutional  
Free: +1 800 868 0441 Home Care  
Fax: +1 630 307 6195  
E-mail: us.info@arjohuntleigh.com

**CANADA**

ArjoHuntleigh  
90 Matheson Boulevard West  
Suite 300  
CA-MISSISSAUGA, ON, L5R 3R3  
Tel/Tél: +1 905 238 7880  
Free: +1 800 665 4831 Institutional  
Free: +1 800 868 0441 Home Care  
Fax: +1 905 238 7881  
E-mail: info.canada@arjohuntleigh.com

**AUSTRALIA**

ArjoHuntleigh Pty Ltd  
78, Forsyth Street  
O'Connor  
AU-6163 Western Australia  
Tel: +61 89337 4111  
Free: +1 800 072 040  
Fax: + 61 89337 9077

**NEW ZEALAND**

ArjoHuntleigh Ltd  
41 Vestey Drive  
Mount Wellington  
NZ-AUCKLAND 1060  
Tel: +64 (0) 9 573 534  
Free Call: 0800 000 151  
Fax: +64 (0) 9 573 5384  
E-mail: nz.info@ArjoHuntleigh.com

# HUNTLEIGH *...performance for life*

## **Huntleigh Healthcare Ltd. - Diagnostic Products Division**

35 Portmanmoor Road, Cardiff, CF24 5HN, United Kingdom

**T:** +44 (0)29 20485885 **F:** +44 (0)29 20492520 **E:** sales@huntleigh-diagnostics.co.uk

**W:** www.huntleigh-diagnostics.com



Registered No: 942245 England & Wales. Registered Office:

ArjoHuntleigh House, Houghton Hall Business Park, Houghton Regis, Bedfordshire, LU5 5XF

©Huntleigh Healthcare Limited 2014

AN **ARJOHUNTLEIGH** COMPANY, MEMBER OF GETINGE GROUP

® and ™ are trademarks of Huntleigh Technology Limited

As our policy is one of continuous improvement, we reserve the right to modify designs without prior notice.